

Professional Manual Cleaning of Laboratory Glassware

Instruction for use	
Product	deconex® 12 BASIC
Concentration	2%
Temperature	room temperature
Time	overnight
Method	soaking bath



Volume of ready-to-use solution	Quantity of deconex® concentrate required		
	1%	1.5%	2%
10 l	100 ml	150 ml	200 ml
20 l	200 ml	300 ml	400 ml
30 l	300 ml	450 ml	600 ml
40 l	400 ml	600 ml	800 ml



Pay attention to the correct dosage!

Recommended procedure



1

Soak the glassware

- Place your glassware in the soaking bath as soon as possible, do not allow residues to dry on the surface. Make sure the items are completely immersed.
- If possible, cover the bath with a lid; follow the recommended soaking time.



2-16 h

2



3

Rinse the glassware thoroughly

- with drinking water
- with demineralized water



4



5

Dry the glassware

- on a drying board or
- in a drying oven



60 °C - 2 h

6

Recommendation: Replace the cleaning bath when it becomes highly contaminated, after the significantly extended cleaning time, or if pH < 9 is reached.